

AMENDMENT

U.S. Appln. No. 09/418,536

PHA 24,285

*Sub B1  
D1*  
7. (Amended) The method of claim 2 wherein prior to the  
replaying step, a user selects which information is replayed.

*Sub B2  
D1*  
9. (Amended) The method of claim 1 wherein the ECG data is  
selected from the group consisting of: patient ECG data and  
patient therapy data.

*B3  
Sub  
D1*  
15. (Amended) The method of claim 2 wherein the replaying  
is optional and the replaying option is presented to a user when  
the instrument is turned off.

16. (Amended) The method of claim 2 wherein the replaying  
is optional and the replaying option is presented to a user when  
a battery is inserted into the defibrillator.

17. (Amended) The method of claim 2 wherein the  
defibrillator continues to monitor the patient's ECG.

AMENDMENT

U.S. Appln. No. 09/418,536

PHA 24,285

B4  
Sub  
C2

19. (Twice Amended) An external defibrillator comprising:  
a controller;  
an energy delivery system operable by the controller to deliver  
an electrical shock from an energy source to an electrode interface;  
memory for recording incident data;  
a screen;  
an incident review activator; and  
an incident review output comprising a visual image  
generator, wherein the incident review output retrieves the  
incident data from memory upon activation of the incident review  
activator by the user and displays the retrieved incident data on  
the defibrillator screen without requiring communication with an  
external device.

REMARKS

Reconsideration and allowance of this application are respectfully requested in light of the above amendments and the following remarks.

The Office Action has been carefully viewed, and Applicants respectfully submit that claims 7,9, 15, 17 and 19 have been amended to overcome rejection under 35 U.S.C. §112, second paragraph by adopting the Examiner's suggested claim language.